

Metals In Mount Isa

Airborne Dust and Metallic Particulates Concentrations


This report must not be reproduced except in full. Results or figures from this report must not be used without acknowledgment




A GLENORE Company

Note for reading these tables: These results show HVAS site performance cumulative average reported over an annual period. The results track our performance towards meeting our Environmental Authority (EA) limit, which is indicated by the red line at the top of the graph.


Cumulative Total Suspended Particulates (TSP)




Cumulative Lead (Pb) in Total Suspended Particulates



Cumulative Arsenic (As) in PM10 Particulates



Cumulative Cadmium (Cd) in PM10 Particulates



Last validated HVAS run day: 04/08/2026

Monitoring Methodology

Analytical Procedures

Total Suspended Particulate (TSP) and Particulate Matter - 10µm (PM10) air quality monitoring is

Metals In Mount Isa



MOUNT ISA
MINES
A GLENORE Company

undertaken using High Volume Air Samplers (HVAS). The HVAS units are operated once every six days for a 24-hour period from midnight to midnight (local standard time – UTC/GMT +10 hours) in accordance with AS/NZS 3580. Volume measurements are corrected to standard temperature and pressure (STP) conditions (0°C, 101.3 kPa). Filter papers are collected between runs and analysed by a NATA accredited laboratory. Analysis and validation of data may take up to four weeks from the

METHOD	ANALYSING LABORATORY	IATA ACCREDITATION
In-house method SW1140522 - Sample Collection	Mount Isa Mines Analytical Laboratory	3978
AS/NZS 3580.9.3 (2015)	Mount Isa Mines Analytical Laboratory	3978
AS/NZS 3580.9.6 (2015)	Mount Isa Mines Analytical Laboratory	3978
In house method EA144 - Metals Analysis on Monitoring Filters	ALS Environmental	825

Calibration and Maintenance

Calibration of the HVAS units is undertaken monthly, in accordance with relevant standards. The table below lists the tolerances used for calibration of the equipment used and the associated

EQUIPMENT	PARAMETER	CALIBRATION TOLERANCE	UNCERTAINTY ¹
Ecotech HiVol 3000 - TSP PM10	Temperature	±1°C of reference meter ²	
	Pressure	±7.5 mmHg of reference meter	
	Pre-Calibration air flow	±10% of calculated value	
	Post-Calibration air flow	±1% of calculated value	±1%
	Time	N/A	Negligible
Cubis Analytical Balance - MSU224S-100-DU	Mass	Per NATA calibration report	±0.07mg

¹Taken from manufacturer specifications

²Reference meters are independently calibrated against a primary source every 6 months by HVAS Manufacturer (Ecotech)

Monitoring Locations

